

RESEARCH

Open Access



Cross-oncopanel study reveals high sensitivity and accuracy with overall analytical performance depending on genomic regions

Binsheng Gong^{1†}, Dan Li^{1†}, Rebecca Kusko^{2†}, Natalia Novorodovskaya³, Yifan Zhang^{1,4}, Shangzi Wang⁵, Carlos Pabón-Peña⁶, Zhihong Zhang⁷, Kevin Lai⁸, Wanshi Cai⁹, Jennifer S. LoCoco¹⁰, Eric Lader¹¹, Todd A. Richmond¹², Vinay K. Mittal¹³, Liang-Chun Liu¹⁴, Donald J. Johann Jr¹⁵, James C. Willey¹⁶, Pierre R. Bushel¹⁷, Ying Yu⁵, Chang Xu¹¹, Guangchun Chen¹⁸, Daniel Burgess¹⁹, Simon Cawley²⁰, Kristina Giorda²¹, Nathan Haseley¹⁰, Fujun Qiu⁷, Katherine Wilkins⁶, Hanane Arib²², Claire Attwooll¹⁰, Kevin Babson²³, Longlong Bao^{24,25,26}, Wenjun Bao²⁷, Anne Bergstrom Lucas⁶, Hunter Best^{28,29}, Ambica Bhandari³⁰, Halil Bisgin³¹, James Blackburn^{32,33}, Thomas M. Blomquist^{34,35}, Lisa Boardman³⁶, Blake Burgher³⁷, Daniel J. Butler³⁸, Chia-Jung Chang³⁹, Alka Chaubey²³, Tao Chen⁴⁰, Marco Chierici⁴¹, Christopher R. Chin³⁸, Devin Close²⁹, Jeffrey Conroy³⁷, Jessica Cooley Coleman²³, Daniel J. Craig⁴², Erin Crawford⁴², Angela del Pozo^{43,44}, Ira W. Deveson^{45,46}, Daniel Duncan⁴⁷, Agda Karina Eterovic⁴⁸, Xiaohui Fan⁴⁹, Jonathan Foox³⁸, Cesare Furlanello^{41,50}, Abhisek Ghosal³⁰, Sean Glenn³⁷, Meijian Guan²⁷, Christine Haag⁵¹, Xinyi Hang⁹, Scott Happe⁵², Brittany Hennigan²³, Jennifer Hipp⁵³, Huixiao Hong¹, Kyle Horvath³⁰, Jianhong Hu⁵⁴, Li-Yuan Hung⁵⁵, Mirna Jarosz⁵⁶, Jennifer Kerkhof⁵⁷, Benjamin Kipp⁵⁸, David Philip Kreil⁵⁹, Paweł Łabaj^{60,61}, Pablo Lapunzina^{62,63,44}, Peng Li⁵⁵, Quan-Zhen Li¹⁸, Weihua Li⁶⁴, Zhiguang Li⁶⁵, Yu Liang⁶⁶, Shaoqing Liu⁶⁷, Zhichao Liu¹, Charles Ma⁴⁷, Narasimha Marella⁴⁷, Rubén Martín-Arenas⁶⁸, Dalila B. Megherbi⁶⁹, Qingchang Meng⁵⁴, Piotr A. Mieczkowski⁷⁰, Tom Morrison⁷¹, Donna Muzny⁵⁴, Baitang Ning¹, Barbara L. Parsons⁴⁰, Cloud P. Paweletz⁷², Mehdi Pirooznia⁷³, Wubin Qu⁹, Amelia Raymond⁷⁴, Paul Rindler²⁹, Rebecca Ringler³⁰, Bekim Sadikovic^{57,75}, Andreas Scherer^{76,44}, Egbert Schulze⁷⁷, Robert Sebra²², Rita Shakhovich⁴⁷, Qiang Shi⁷⁸, Tielu Shi⁷⁹, Juan Carlos Silla-Castro⁸⁰, Melissa Smith²², Mario Solís López^{43,44}, Ping Song⁴⁸, Daniel Stetson⁷⁴, Maya Strahl²², Alan Stuart⁵⁷, Julianna Supplee⁷², Philippe Szankasi²⁹, Haowen Tan⁸¹, Lin-ya Tang⁴⁸, Yonghui Tao^{24,25,26}, Shraddha Thakkar¹, Danielle Thierry-Mieg⁸², Jean Thierry-Mieg⁸², Venkat J. Thodima⁴⁷, David Thomas^{45,33}, Boris Tichý^{83,44}, Nikola Tom^{83,44}, Elena Vallespin Garcia^{43,44}, Suman Verma³⁰, Kimbley Walker⁵⁴, Charles Wang^{84,85}, Junwen Wang^{86,87,88}, Yexun Wang¹¹, Zhining Wen⁸⁹, Valtteri Wirta⁹⁰, Leihong Wu¹, Chunlin Xiao⁹¹, Wenzhong Xiao^{55,39}, Shibe Xu⁹², Mary Yang⁴, Jianming Ying⁶⁴, Shun H. Yip^{86,93}, Guangliang Zhang⁹⁴, Sa Zhang⁹⁴, Meiru Zhao⁹⁵, Yuanting Zheng⁵, Xiaoyan Zhou^{24,25,26}, Christopher E. Mason³⁸, Timothy Mercer^{96,97}, Weida Tong¹, Leming Shi^{5,98,99*}, Wendell Jones^{100*} and Joshua Xu^{1*} 

* Correspondence: lemingshi@fudan.edu.cn; wendell.jones@q2labsolutions.com; joshua.xu@fda.hhs.gov

[†]Binsheng Gong, Dan Li and Rebecca Kusko contributed equally to this work.

⁵State Key Laboratory of Genetic Engineering, School of Life Sciences and Shanghai Cancer Hospital/Cancer Institute, Fudan University, Shanghai 200438, China

¹⁰⁰Q2 Solutions - EA Genomics, 5927 S Miami Blvd, Morrisville, NC 27560, USA

¹Division of Bioinformatics and Biostatistics, National Center for Toxicological Research, US Food and Drug Administration, Jefferson, AR 72079, USA

Full list of author information is available at the end of the article



© The Author(s). 2021, corrected publication 2021. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

Abstract

Background: Targeted sequencing using oncopanels requires comprehensive assessments of accuracy and detection sensitivity to ensure analytical validity. By employing reference materials characterized by the U.S. Food and Drug Administration-led SEquence Quality Control project phase2 (SEQC2) effort, we perform a cross-platform multi-lab evaluation of eight Pan-Cancer panels to assess best practices for oncopanel sequencing.

Results: All panels demonstrate high sensitivity across targeted high-confidence coding regions and variant types for the variants previously verified to have variant allele frequency (VAF) in the 5–20% range. Sensitivity is reduced by utilizing VAF thresholds due to inherent variability in VAF measurements. Enforcing a VAF threshold for reporting has a positive impact on reducing false positive calls. Importantly, the false positive rate is found to be significantly higher outside the high-confidence coding regions, resulting in lower reproducibility. Thus, region restriction and VAF thresholds lead to low relative technical variability in estimating promising biomarkers and tumor mutational burden.

Conclusion: This comprehensive study provides actionable guidelines for oncopanel sequencing and clear evidence that supports a simplified approach to assess the analytical performance of oncopanels. It will facilitate the rapid implementation, validation, and quality control of oncopanels in clinical use.

Keywords: Oncopanel sequencing, Target enrichment, Molecular diagnostics, Reproducibility, Analytical performance, Precision medicine

Background

Despite ongoing advances in sequencing technologies to date, the promise of their use for precision medicine has not yet borne out entirely for the majority of cancer patients. Although cancer is a genomic disease, most malignant tumors are currently classified and often treated based on the tissue of origin and the TNM (tumor, node, and metastases) staging system for solid tumors rather than presence or absence of relevant mutations in key oncogenic pathways. Presently, only a small number of specific mutations in genes such as *EGFR*, *ALK* [1], and *BRAF* [2] are routinely tested to inform clinical decisions. To move precision medicine from theoretical discussions to the bedside of individual patients, it requires accurate and reproducible testing of a broad spectrum of clinically relevant mutations combined with appropriate therapies. Hence, improved mutation testing is greatly needed to confidently classify individual tumors, reliably provide personalized therapeutic choices, and inform patient prognosis.

Next-generation sequencing (NGS) of targeted genomic regions has become increasingly common because targeted panels focus sequencing coverage on genomic regions more likely to harbor important genomic aberrations and thereby increase diagnostic sensitivity [3]. Oncopanels have several potential clinical applications, including identification of driver and potential driver mutations. Variants can be prioritized for their biological significance and cross-referenced with potentially useful targeted therapies. Some oncopanels are also used to assess micro-satellite instability [4], homologous recombination defect status [5], copy number alterations [6], and gene rearrangements [7]. There is increasing evidence that a significant proportion of patients with low-frequency alleles of these clinically actionable mutations show clinical response to targeted therapies [8]. Because key therapeutic decisions are based on these results,

including choice of molecular-targeted therapy and prediction of outcomes, it is critical that mutation detection displays high sensitivity and low false positive (FP) rate and that reports to physicians are highly accurate [9–11]. Importantly, both oncopanels and downstream bioinformatic methods must be validated to be accurate, informative, and reliable.

Certain tumors that are genetically unstable may result in a large number of cellular proteins having somatic mutations (also known as neoantigens). More exposure of neoantigens increases the likelihood of non-self recognition by the adaptive immune system. If such a recognition event has occurred, but the tumor still progresses, a possible reason is the autoregulatory process of immune checkpoints which ordinarily prevent autoimmunity by deactivating active T cells in the tumor microenvironment. Therefore, if oncopanels are large enough in size, they can be used to estimate tumor mutational burden (TMB), an emerging indicator for immune checkpoint inhibitors in certain tumor types [12, 13].

In current practice, there are multiple Clinical Laboratory Improvement Amendments (CLIA)-certified labs that have implemented oncopanels for tumor profiling. At present, four pan-cancer panels have been approved or cleared by U.S. Food and Drug Administration (FDA): FoundationOne CDx [14], MSK-IMPACT [10], NantHealth Omics Core [15], and PGDx elio tissue complete [16]. In 2017, the FDA approved Merck's Keytruda (pembrolizumab) for any unresectable or metastatic solid tumor with evidence of mismatch repair deficiency. Thus, this was the first approval that was biomarker-based and tissue agnostic [17]. The FDA accelerated the approval of Bayer's Vitrakvi in 2018, and Roche's Rozlytrek (entrectinib) in 2019, both of which are biomarker-based cancer treatments [18]. However, only a minority of cancer patients undergo high-throughput mutational profiling [19] and many physicians remain skeptical of the rigorousness of genomic testing [20], hesitant to deplete precious tumor material, and/or apprehensive about performing an additional tumor biopsy [21]. For oncopanel sequencing to become a useful and widely adopted clinical test, a thorough understanding of its sensitivity, reproducibility, and accuracy is greatly needed. To date, no study has assessed how distinct oncopanel methodologies perform with respect to variant allele frequency (VAF) and variant location, impacting sensitivity, FP rate, reproducibility, and potential biomarkers such as TMB for oncopanels. Previous studies comparing oncopanels have not used reference materials as test samples with (1) enough VAF range and sufficient diversity of mutations to thoroughly interrogate a sufficient number of clinically relevant mutations [22] and (2) sufficient material readily available for the scientific community [23–25]. In contrast, this study used recently characterized reference samples that contain an extensive collection of known positive and negative variants [26] to provide a much more definitive and comprehensive understanding of key variables that impact performance of oncopanels.

The FDA-led SEquencing and Quality Control project phase 2 (SEQC2) continues earlier consortium efforts [27–30] to advance best practices and address the technical gaps in genomics technologies. Its Somatic Mutation Working Group has extensively identified variant content of an individual pair of reference samples (breast cancer cell line HCC1395 and its normal counterpart) and established best practices for genome-wide somatic mutation calling [31, 32]. In parallel, the Oncopanel Sequencing Working Group's efforts, which are detailed herein, utilize a recently identified novel set of

comprehensive variants and negatives [26] from additional reference samples to compare variant calling across eight distinct oncopanel. We assessed intra-lab and cross-lab reproducibility, sensitivity for known positives, and accuracy (i.e., FP rate) through known negatives across a range of VAFs and in selected important genomic regions. The relationships between analytical variables and TMB estimation were also established to inform the selection of oncopanel and harmonization of TMB measurements across test platforms, which are important needs [33] identified recently to advance the successful utilization of TMB in clinical practice. This study provided a comprehensive evaluation of eight oncopanel to inform best practice guidelines for oncopanel sequencing (Table 1).


Results

Overview of study design

We employed four reference samples (Fig. 1a)—sample A (an equal mass pooled genomic DNA (gDNA) sample of ten cancer cell lines from the Agilent Universal Human Reference RNA (UHRR) material) [34], sample B (derived from a non-cancer male cell line), sample C (equal dilution of sample A and sample B), and a spike-in sample (sample B with 5% AcroMetrix hotspot synthetic controls [35]). The majority (about two

Table 1 Summary of recommendations

Issue	Recommendation
Reference material selection	A reference material (or a set of reference materials) with a high density of known variants spanning a range of low allele frequencies (e.g., from above 1 to 20%) is needed to assess the analytical performance of oncopanel (Fig. 1a). Ideally, the reference material will encompass a diversity of simple and complex variants in different cancer-associated genes.
Sensitivity	Sensitivity was found to be high (> 96.5%) for variants previously verified to have variant allele frequency (VAF) greater than 5% (Fig. 2a). We recommend that a less stringent VAF threshold is used to achieve high sensitivity for variants of VAF below 5%.
Spike-in	Utilizing a sample spiked-in at a specified amount (e.g., 5%) can provide additional variants at known allele frequencies for analytical validation of oncopanel.
Variant type	The sensitivity for detecting insertion and deletion variants (indels) is typically more variable and poorer than single nucleotide variants (SNVs), and this difference becomes more pronounced at low VAFs (Fig. 3). We recommend that analytical validation of indels is performed independently to SNVs.
Determining VAF threshold	Reference materials can be used to establish an optimal VAF threshold that reduces false positives (FPs) and retains sensitivity. Indels and SNVs may require different VAF thresholds to optimize performance (Fig. 3b).
Controlling the FP rate	In applications where a minimal FP rate is required, raising the allele frequency threshold was effective at reducing FPs. The additional restriction of analysis to the consensus targeted regions (CTR) can further reduce FP rate.
Genomic location	Genomic location can impact the rate of FPs detected, and we recommend that analytical validation of panels is independently performed inside and outside of the CTR (Fig. 4a).
Cross-lab reproducibility	Measuring the analytical performance of a panel in multiple labs is critical to establish reproducibility (Fig. 4c).
Estimating tumor mutational burden (TMB)	TMB estimation should be confined to the CTR of each panel. Applying a minimal VAF threshold was helpful to reduce FPs and improve TMB evaluation (Fig. 5d).



thirds) of more than 42,000 known small variants (including SNVs, small indels, and multi-nucleotide variants (MNVs)) in the 22-Mb exon coding regions of sample A have a VAF below 20%, making sample A uniquely suitable for assessing the analytical performance of oncopanels. Sample C was included to increase the number of known variants with VAF between 1 and 2.5%. More details about the four reference samples are included in “Methods.” A total of eight oncopanels from eight distinct well-established panel providers were tested herein on these four samples, including Agilent Custom Comprehensive Cancer Panel v2 (AGL), Burning Rock DX OncoScreen Plus (BRP), Integrated DNA Technologies xGen Pan-Cancer Panel (IDT), iGeneTech AIOnco-seq (IGT), Illumina TruSight Tumor 170 (ILM), QIAGEN Comprehensive cancer panel (QGN), Roche SeqCap EZ Choice custom PHC Panel (ROC), and Thermo Fisher Oncomine Comprehensive Assay v3 (TFS). Vendors shipped the same reagent kits to all sites. Information about the oncopanels is included in Fig. 1b and Additional file 1: Table S1.

Each panel was tested in at least three independent laboratories with quadruplicates for each sample. In total, this unprecedented large-scale study encompassed 28 laboratories and 430 libraries (Fig. 1c). Each panel provider recruited the test laboratories and harmonized the experimental protocols to minimize cross-lab variations in the study. Libraries were sequenced on the same sequencing platform for each panel and the panel provider then performed panel-specific variant calling. For more information about sequencing and downstream bioinformatic pipelines, please see Additional file 2: Supplementary Methods. Variant calling results were submitted in variant call format (VCF) files, along with their targeted regions and whitelist/blacklist, to an independent group with sole knowledge of the known content of samples A, B, C, and the spike-in for performance evaluation (Fig. 1d). Intra-lab and cross-lab reproducibility were

computed between library replicates as a measure of variant calling consistency. Sensitivity and FP rate were then estimated by utilizing the large sets of known positive (KP) and known negative (KN) loci from our companion work [26]. More than 42,000 KPs and 10 million KNs in sample A were characterized in the 22-Mb high-confidence exon coding regions through extensive sequencing of sample A and the individual cell line DNA samples that made up sample A. KN loci had the mismatch error rate below 0.25%. The distribution of KPs by VAF is shown in Fig. 1a.

Sensitive detection of SNVs in the focused regions by all panels

We evaluated the performance of detecting SNVs in the consensus targeted regions (CTR, see “Methods” for details) [26]. Given there is a general consensus that higher VAF variants are “easier” to detect, we focused our analysis on metrics that characterized performance relative to VAF, especially in the lower [1–10%] VAF range (Fig. 2). The number of KPs and observed sensitivity are listed side by side for each VAF range across three reference samples (A, C, and spike-in) and all panels (Fig. 2a). Sample C was not tested by TFS. Overall, sensitivity across all VAF ranges in all panels was high with narrow confidence intervals (Additional file 1: Table S2). It ranged from 87.1 to 98.3% for the lowest VAF range (1–2.5%) tested by AGL, BRP, IGT, and QGN (Fig. 2a). Sensitivity was greater than 96.8% for the 2.5–5% range and greater than 98.4% for the 5–20% range. The unusual sensitivity for ILM in sample C for the 10–20% range was due to a single false negative in all library replicates. Other exceptions were ILM, ROC, and TFS for the 2.5–5% range, as their variant calling pipelines had a VAF reporting threshold very close to 2.5%. So their sensitivity was reduced due to a boundary VAF threshold effect (see below for the effect of VAF cutoff on sensitivity). Accordingly,




Fig. 2 Reproducibility and sensitivity across VAF ranges for SNVs in the consensus targeted regions. **a** Table listed the number of known variants in each VAF range (left number), sensitivity (right number) for all 8 panels across all samples tested. For the panels with a built-in VAF threshold, “N/A” is listed if the VAF low bound is much lower than the panel provider’s chosen VAF threshold. The VAF threshold is 2.6% for ILM, 2.0% for IDT, 2.5% for ROC, and 2.5% for TFS, respectively. **b** Average false positive SNV calls per million across various VAF cutoffs. Jittering was applied to avoid overlapping. **c** Cross-lab and intra-lab reproducibility (in Phred scale) for variant calls with VAF between 2.5 and 20%


these threshold-dependent panels and IDT were not assessed for the lowest VAF range (since IDT's pipeline utilizes a VAF reporting threshold of 2.0%). All false negatives in sample A or C were manually checked and can be explained by a combination of low VAF and the panel's VAF threshold except the above mentioned single case for ILM in sample C. We initially anticipated similar sensitivity for the spike-in sample in the VAF range of 2.5–10%, as the variants were designed to be 5% VAF. However, this was not observed, likely due to interference of the nearby spike-in variants with amplicon primers, DNA capture, or variant calling. To derive sensitivity estimates, custom filtering of the spike-in's variants was carried out by the panel providers to exclude variants under primer interference for QGN and TFS or variants in dense clusters for IDT (see Additional file 2: Supplementary Methods for details). Likewise, several dense variant clusters led to unexpectedly low reported sensitivity for ROC (88.0%).

All panels had FP rates lower than 10.5 per Mb in the CTR with each panel's default VAF threshold, and the FP rates dropped as the VAF cutoff increased (Fig. 2b). The FP rate was noticeably different between panels for a VAF cutoff at 1%, but became increasingly similar as the VAF cutoff became more stringent. As a result, all panels had similar FP rates of approximately 1 FP per Mb or less for VAF greater than 5%. And all panels reported no FP calls for VAF greater than 10%. We also assessed both intra- and cross-lab reproducibility for variant calls in the CTR with VAF between 2.5 and 20% and observed that they were similar for each panel sample combination (Fig. 2c, Additional file 3: Fig. S1A), with cross-lab reproducibility being slightly worse than intra-lab reproducibility in a few instances. Results for sample A and sample C were similar for VAF above 5% but showed some noticeable differences in the VAF range 2.5–5%. The differences may be driven by VAF reporting boundary effects, as sample C has more variants in the lower VAF ranges (thus, near the VAF reporting thresholds for several panels) (Additional file 3: Fig. S1B).

Impact of variant type, VAF cutoff, and genomic region on sensitivity

We explored the impact of variant type on sensitivity for each panel by aggregating all sample A and C libraries (Fig. 3a). Non-SNV variant calling (small indels and multi-nucleotide variants (MNVs)) had more variable sensitivity than SNV variant calling for AGL, ILM, and ROC panels, as the number of known non-SNV small variants per panel in samples A and C is typically small by absolute numbers (green numbers in Fig. 3a, Additional file 3: Fig. S2). Thus the confidence interval was usually wider for non-SNV small variants (Additional file 1: Table S3). We characterized the impact of VAF on sensitivity comparing SNVs against non-SNV variants (Additional file 3: Fig. S2). For sample A, the difference between SNVs and non-SNV variants was most pronounced in the lowest VAF range. For both SNVs and non-SNV variants, sensitivity was generally observed to improve with known variants of higher expected VAFs.

We also applied a range of artificial VAF cutoffs in addition to each panel's normal reporting thresholds from stringent (2.5%) to less stringent (1.5%) to better understand the impact of a VAF reporting threshold on sensitivity in detecting known positives in the VAF range 2.5–5% (Fig. 3b, see Additional file 1: Table S4 for confidence intervals). As the cutoff filter was more relaxed, and the boundary effects decreased, sensitivity rose expectedly across most panels and all reference samples. This increased sensitivity



was more prominent in samples A and C, which may be explained by the fixed and higher VAF (5%) for variants in the spike-in sample. The boundary effects became negligible for the higher VAF ranges, i.e., 5–10% and 10–20%, in samples A and C (Additional file 1: Table S5).

Sensitivity within the CTR and outside of the CTR, but inside the High Confidence Coding Region (HC_CR, see “Methods” for details), was compared (Fig. 3c). For samples A and C, most panels performed consistently within and outside of the CTR (but within HC_CR). TFS, ILM (for sample C), and ROC had higher observed sensitivity outside of the CTR, but also had relatively few KPs in this region (Fig. 3c top bar graph). Thus narrower confidence intervals for sensitivity were achieved within the CTR (Additional file 1: Table S6).

False positive rates were significantly higher beyond the focused regions

We employed three different methods to estimate the FP rate. Method 1 relies on the KNs (“via_KN”) in samples A and C (Additional file 3: Fig. S3A). Our list of KNs is extensive (> 10 M) but only available within the CTR (which is the majority of the coding region) and comprises about 30% of all coding regions. Using Method 1, a FP is a variant call on a KN position. A stringent consensus across multiple sequencing experiments was employed to determine the partial list of negatives within the CTR [26]. Method 1 may tend to underestimate the overall FP rate across all positions of the CTR. Method 2 marked any call in the normal sample B with VAF between 1 and 10% as a false positive (“B_low,” Additional file 3: Fig. S3B) due to strong evidence of this sample having normal diploid cell characteristics [26]. Method 3 leveraged sample mixture to differentiate FPs. Any variant call in sample C was deemed an FP if it was not called in any of four sample B libraries with VAF above 10% or any of four sample A libraries at the same test laboratory (“C_only,” Additional file 3: Fig. S3C). Method 2

might overestimate the FP rate by failing to exclude rare mosaic variants caused by cell line drift. Method 3 could underestimate the FP rate by failing to recognize those FP calls due to systematic errors as they would be called in all three samples. These three methods were compared and found to produce largely similar estimates within the CTR (Additional file 3: Fig. S3D), with the “B_low” method in one specific VAF range of IGT and ILM being the only exception. As expected, almost all FP calls within the CTR were irreproducible across library replicates for any panel, sample, and FP method (Additional file 3: Fig. S4). Those few exceptions were then confirmed through manual curation as systematic errors (see the legend for Additional file 3: Fig. S4).

Using the last two methods (“B_low” and “C_only”), we compared the FP rate within and outside of the CTR (Fig. 4a). By either method, across VAF ranges and across panels, we observed that the FP rate was lower in the CTR than outside. Breaking each panel into three sub-regions (inside the CTR, outside the CTR but in HC_CR, or the rest with the subregion sizes listed in Additional file 3: Fig. S3E), the FP rate was the lowest in the CTR and the highest outside of the HC_CR when using either the B_low or C_only method (Additional file 3: Fig. S3F-G). FP rates inside the CTR across different VAF cutoffs are listed in Fig. 4b. For all panels, the FP rate decreased as the VAF increased.

Finally, to better understand the impact of genomic regions on cross-lab reproducibility, we assessed cross-lab reproducibility within and outside of the CTR with VAF between 2.5 and 20% (Fig. 4c). Across most panels and in both samples A and C, cross-lab reproducibility was marginally lower outside of the CTR. Taken together, these results suggest that variant location may drive high cross-lab reproducibility seen in





Fig. 4 Impact of CTR region on FP rate and reproducibility. **a** FP rate in and outside of the CTR using two different methods (B_low and C_only) at three different VAF cutoffs, 1%, 2.5%, and 5%. C_only was not applied to TFS as sample C was not tested on TFS. FP rates are plotted in squared root scale. **b** Estimated FP rate within the CTR averaged over three methods at different VAF cutoffs. **c** Cross-lab reproducibility (in Phred scale) in samples A and C within and outside of the CTR

Fig. 2c. Since sensitivity did not drop noticeably outside the CTR (Fig. 3c), the observed drops in reproducibility (Fig. 4c) may be driven by more FPs outside of the CTR (Fig. 4a).

Enabling oncopanels of sufficient size to estimate TMB

Commonly measured by whole-exome sequencing (WES), TMB ranges over several orders of magnitude and varies by tumor type. Suitable TMB biomarker thresholds can also vary by tumor type, implying different degrees of accuracy may be required for different indications. The elevated FP rate outside of the CTR will have a strong detrimental effect on the accuracy of variant calls and subsequently TMB estimates. This leads to a reasonable requirement that TMB estimation should be confined to the CTR portion of each panel. The high density of variants in these reference samples was purposely designed to enable an accurate evaluation of sensitivity and limit of detection. To investigate a more “real-world” scenario where using an oncopanel to estimate exome-level TMB might be relevant, we randomly selected known variants comprised of a prevalence more consistent with actual tumors for a variety of indications (e.g., between 5 and 50 per Mb). We then computed the average technical run-to-run variance of panel-focused TMB estimates and plotted the technical coefficient of variation (CV) at two different VAF cutoffs:



2.5% (Fig. 5a) and 5% (Fig. 5b, see “Methods” for details). All six panels with targeted regions in the CTR greater than 250 kb were included in the analysis. Clearly, the technical CV decreased at higher panel TMB values. There was no observed effect related to the panel size. All panels achieved a CV lower than 25% when panel-focused TMB was 25 or higher.

To assess the magnitude of additional variance introduced by the genomic coverage of oncopanels, we utilized The Cancer Genome Atlas (TCGA) WES mutation data [36] to compute TMB estimates using each panel’s CTR portion. We then used the mean squared deviation (MSD) to measure the TMB estimation difference between whole-exome and the panel-specific CTR portion. MSD has two components: mean bias and variance. We characterized an intrinsic CV measure specific to each panel, related to this variance component. Further linear regression analysis led to an approximate equation estimating the intrinsic CV by TMB and the panel size (Additional file 3: Fig. S5, see “Methods” for details). The intrinsic CV is plotted in Fig. 5c. In agreement with technical CV, intrinsic CV decreases at higher TMB values. For example, a 1-Mb oncopanel is estimated to have an intrinsic CV of 25% at a TMB value of 21. By incorporating both technical variance and intrinsic variance, an overall CV for TMB was computed and plotted in Fig. 5d when using VAF cutoffs of 2.5% and 5% for each panel. Overall CV was close to intrinsic CV for each panel at corresponding TMB values. This indicates that all panels were able to achieve a lower technical variance than the intrinsic variance associated with estimating exome-level TMB from a smaller focused panel. Nevertheless, lower technical run-to-run variance is possible for all panels using a higher VAF cutoff as illustrated in Fig. 5d. This improvement was more apparent for AGL and QGN panels at lower TMB, and consistently large for IDT across a wide TMB range. Although many of the panels, e.g., QGN, were not designed for TMB estimation, they may be used to do so under certain circumstances. However, if the goal is to maintain an overall CV lower than 25% in the clinically important TMB range between 15 and 30, AGL combined with an VAF cutoff at 5% would be the only panel from these eight suitable for TMB estimation.

Discussion

This oncopanel study was comprehensive in several areas: (1) the use of rigorous performance metrics both within and outside of easier/difficult genomic regions, (2) utilization of various oncopanel vendors with different panel sizes, (3) broad multi-laboratory participation, (4) analysis of a wide dynamic range of VAFs, and (5) use of test samples comprising advanced reference material with dense, known variant content from a stable source. We assessed reproducibility, sensitivity, and the FP rate of eight pan-cancer oncopanels across laboratories. We examined the impact on these metrics by VAF ranges, variant types, and genomic regions. These panels ranged in size from as small as 149 up to 7625 kb, and DNA input ranged from 20 to 100 ng. Enrichment of targets by both hybridization-based capture and PCR amplification were represented. The panel vendors represented major providers of currently marketed targeted sequencing technologies.

We cannot be completely certain that all of our results would generalize to a new admixture reference sample similarly constructed or real-world clinical formalin-fixed paraffin-embedded (FFPE) samples. Our primary reference sample (sample A) does not

reflect all cancer indications, nor does it reflect all subtypes for any given cancer [26]. However, this approach provided noticeably better estimates than previous efforts [24]. For example, the large number of variants present in diverse sequence contexts in samples A and C enabled an accurate assessment of sensitivity and reproducibility for detecting SNVs. We observed that these panels detected known small indel or MNV positives with similar sensitivity and reproducibility as SNVs (Fig. 3a, Additional file 3: Fig. S2). However, with the limited amount of small indel/MNV positives identified as a proportion of all identified variants in the reference sample (1.2%), it is likely given prior research in the field that these small indel/MNV positives identified in the reference sample are in group that is easier to detect, in general. Therefore, we cannot suggest that these panels find small indels/MNVs with the same proficiency as SNVs throughout their target regions until we have identified a larger and more representative set of indels/MNVs from which to compare. While SNVs predominate clinically relevant oncology mutations, non-SNVs can also be key drivers [37], and panels should be further characterized for analysis of non-SNVs as these variants are typically more challenging to detect than SNVs. Additional reference material with more complex variants is required to address this outstanding question.

We chose to use the output of vendor-implemented bioinformatics pipelines for their respective proprietary panels. Other SEQC2 work has clearly demonstrated that choices of bioinformatics pipelines will influence the variant calling results in WES and whole-genome sequencing (WGS) [31]. Our raw data are made public to enable further evaluation of various bioinformatics pipelines by the community. This study did not characterize the effect of read coverage, which can be critical for low VAF variants. Low read coverage can reduce sensitivity and sometimes inflate the FP rate [38]. In addition, this study did not explore the impact of sample quality on variant calling. In the clinical setting, it can take considerable time before samples are processed, and samples are often subject to formalin fixation. A separate team from the OncoPanel Sequencing Working Group revealed the impact of formalin fixation time and within block position on the performance of oncopanel sequencing [39]. Taken together, the experience gained from this study suggests that future analysis should focus on assessing the impact of bioinformatics pipelines, including variant callers, on sensitivity and FP rates.

The FP rate impacts TMB and overall confidence in the panel. To the best of our knowledge, this is the first study with a comprehensive analysis of the FP rate for oncopanels. Importantly, our results revealed that increasing the VAF reporting threshold reduced the FP rate across all VAF values within the CTR, but this effect was far less dramatic outside of the CTR. Therefore, one straightforward way to control the FP rate, at least within the CTR, is to adjust the VAF reporting threshold, which for different panel implementation modalities, may be optimally and more simply determined by assaying our reference samples and additional negative (normal) reference material using multiple technical replicates and cross-lab testing. On the contrary, if the goal is to have sensitive detection of clinically actionable mutations, where the FP rate over a region is not so relevant, lower variant-specific VAF threshold can be used without compromising accuracy through a novel synthetic spike-in control that directly measures the variant-specific and experiment-specific background errors. The control is presented in an accompanying SEQC2 manuscript [40].

To understand the impact of genomic region on sensitivity, reproducibility, and FP rates, we assessed all variants compared to those only within the CTR. Sensitivity was consistent for each panel between the CTR and HC_CR (Fig. 3c). However, because known variants were only available within HC_CR, we were not able to assess sensitivity for the regions outside of HC_CR. We hypothesize that sensitivity would be the same outside of HC_CR or slightly lowered. As for the FP rate, we observed multiple fold increases for all panels outside of the CTR when studying the impact of genomic region (Fig. 4a). Reproducibility was lower outside of the CTR (Fig. 4c). A plausible inference would be that the much higher FP rate outside of the CTR led to lower reproducibility in that region. “Real-world” clinical samples likely harbor much fewer low VAF variants (i.e., VAF < 20%) than sample A or C that was used here. With fewer variants, the effect of FP rate on reproducibility will be even more pronounced, i.e., the change in reproducibility will be greater between the CTR and outside of the CTR. In the extreme case of samples without any low VAF variants, e.g., a pure sample from a healthy infant, reproducibility will be completely dependent on the FP calls with no relationship to sensitivity. When reproducibility would play an important role in the specific application of oncopanel sequencing, for example, profiling multiple samples to study intratumor genetic heterogeneity, one may consider restricting the analysis to the CTR. Although most pathogenic mutations are within the CTR [26], restricting the analysis of clinical samples to the CTR may lead to the exclusion of certain mutations of interest. Currently, the CTR has excluded difficult regions and places likely to cause systematic sequencing errors. Extending the CTR will be a good direction for future work.

We noted that cross-lab reproducibility is similar to, but generally lower than intra-lab reproducibility (Additional file 3: Fig. S1A). The VAF range has a greater impact than the lab on reproducibility; however, given the increased variability across labs, we encourage cross-lab testing of oncopanels. If this is not possible, an alternative is more extensive testing within the same lab with different operators over an extended period. Furthermore, routine proficiency testing will help safeguard the performance, even within the same lab.

When screening clinically actionable mutations, FP variants are typically filtered by various processes, including variant prioritization (where only selected variants are considered for determining targeted therapy), and therefore have a small impact in final clinical decisions [10, 25]. In contrast, FP calls could easily confound TMB calculations and create a noninformative biomarker. Based on the findings of this large-scale cross-panel study, we propose different filtering and threshold-related methods for different modalities: variants for prioritization and targeted therapy selection may employ lower VAF reporting thresholds to achieve maximal sensitivity, while variants used in TMB calculations for immunotherapy purposes should consider higher VAF reporting thresholds or use VAF cutoffs commensurate with a low FP rate. Within this study, we have evaluated the impact of analytical variables on the detection of mutations and TMB estimation without reference to specific cancer subtypes. While the described relationships between analytical variables and TMB estimation are likely to remain constant regardless of cancer subtypes, the cancer type may inform the selection of oncopanel and VAF threshold to optimize diagnostic yields. For example, a greater mutation density, which is typical in lung cancers and melanomas, will enable a more confident estimation of TMB with lower overall CVs (as shown in Fig. 5d for high

TMB values). Therefore, a moderately sized oncopanel with a higher VAF threshold may be sufficient. Alternatively, lower VAFs, which confound confident diagnosis, are typical in cancers with lower sample purity or greater intratumor heterogeneity, such as pancreatic cancers. In these cancers, a larger oncopanel with a lower VAF threshold may be required to confidently estimate TMB.

In summary, all eight panels achieved high sensitivity (ranging from 84.2 to 100%) for variants of low VAF between 1 and 5% (Fig. 2a). Applying an increasingly strict VAF cutoff reduces the FP rate, but also reduces sensitivity (Fig. 3b). VAF reporting thresholds create boundary effects due to stochastic sampling, where a variant may have an observed VAF just above the threshold in one replicate and but below in another, causing a decrease in sensitivity and reproducibility. When comparing performance metrics in different genomic regions, sensitivity remains consistent; however, the FP rate is higher and consequentially reproducibility is lower outside of the CTR. These oncopanels demonstrated low relative technical variability in estimating TMB; however, due to the greater variability in sampling a much smaller region of the exome (Fig. 5), our results suggest that, at a minimum, oncopanels will require 1 Mb in the CTR region to properly estimate TMB at levels commonly used for decision thresholds and may require even larger panels if TMB thresholds are needed below 10–15 mutations per megabase. Results are generally consistent with previous *in silico* studies [41–43]. This comprehensive study provides actionable guidelines for oncopanel sequencing and clear evidence that supports a simplified approach to assess the analytical performance of oncopanels. It will facilitate the rapid implementation, validation, and quality control of oncopanels in clinical use.

Methods

Testing samples

Realizing that there is no adequate reference genomic material to enable a transparent cross-lab study of oncopanels, the Oncopanel Sequencing Working Group developed and validated a genomic sample suitable for benchmarking oncopanel's performance in detecting small variants of low allele frequency (AF). The design, validation process, and the results are reported in a companion paper [26]. This sample, termed "sample A," is composed of an equal mass pooling of 10 gDNA samples prepared from Agilent's Universal Human Reference RNA (UHRR) [34] cancer cell lines. Over 42K small variants are called with high confidence in the defined CTR of over 22 million bases. Briefly, CTR is the overlap of targeted regions of four whole-exome sequencing panels, UCSC coding genes, Ensemble exons, and NIST high-confidence genomic regions, followed by a removal of low complexity regions [26]. The majority of these small variants have a VAF below 20%, making sample A uniquely suitable for assessing the analytical performance of oncology panels. More specifically, about 7% of them are in variant allele frequency (VAF) range (1%, 2.5%), 18% in (2.5%, 5%), 25% in (5%, 10%), and 18% in (10%, 20%). Sample B is a gDNA sample from a normal male cell line (Agilent Human Reference DNA, Male, Agilent part #: 5190-8848). Sample C is a 1:1 mix of sample A and sample B. Sample C was included to increase the number of known variants with VAF between 1 and 2.5%. The spike-in sample is sample B with 5% Acro-Metrix²⁹ Spike hotspot synthetic controls (Thermo Fisher Scientific, Fremont, CA).

Samples A, B, and C were prepared by Agilent and provided without charge to SEQC2 for study. The spike-in sample was prepared at Thermo Fisher Scientific (Fremont, CA) after receiving sample B from Agilent under their material transfer agreement. All four samples were stored in low-EDTA TE buffer (10 mM Tris, 0.1 mM EDTA, pH 8.0) at 20 ng/ μ l concentration.

Participating panels, test sites, and study protocol

Study participation of oncopanel vendors was actively sought out through all available venues. A study plan was discussed in detail with each interested contact, emphasizing the cross-lab study design, transparency, and the requirement for eventual data release to the public domain. Eight panel providers agreed to participate. Each panel provider then recruited independent test sites (i.e., laboratories) that were proficient with its panel and oncology sequencing procedure. Test samples were distributed to test sites after executing Agilent material transfer agreement and SEQC2 confidentiality agreement. A sample processing and sequence data reporting standard operating procedure (SOP) was also distributed to all test sites. It includes test sample description, sample receiving and preparation, a list of test sites, naming convention, testing procedure, and instructions on library preparation quality data collection. There were at least three independent test sites for each participating panel. Panel providers and their supporting partners distributed panels and reagents to the corresponding test sites. Each site followed the SOP and a panel-specific experimental protocol (see Additional file 2: Supplementary Methods for details) to make 4 replicate libraries for each sample. Sequencing was done on the same sequencing platform for each panel. Each panel provider determined the depth of coverage per library and sample C libraries were sequenced twice as deep as other samples. Panel providers were encouraged to conduct pilot sequencing experiments with test samples at their internal lab or test sites. Pilot data was not required for submission to SEQC2.

Reproducibility by VAF ranges per panel

Given the variability in VAF measurements and its influence on reproducibility, a non-symmetric reproducibility measure was devised. For comparing library “*Libx*” to another replicate “*Liby*” of the same sample, reproducibility is defined as the portion of *Libx*’s variant calls that are also detected in *Liby*. It can be computed by dividing the number of variants called in both *Libx* and *Liby* by the number of variant calls in *Libx*. Similarly, a reproducibility value can be computed for comparing *Liby* to *Libx*. These two reproducibility values are not symmetric and thus may be different. For comparing *Libx* to *Liby*, VAF ranges are applied unsymmetrically to compute reproducibility through stratifying variant calls of *Libx* into VAF ranges with no restriction imposed to variant calls of *Liby*. When both library replicates are from a same test lab, it is counted as one intra-lab reproducibility measurement. When they are from different test labs, it is counted as one cross-lab reproducibility measurement. As there are multiple library replicates from multiple labs, many reproducibility measurements can be computed for each sample. The average intra-lab reproducibility indicates what percentage of variants called by one library in one lab within a specific VAF range will be detected by another

library replicate in the same lab without any restriction to their VAF values. The average cross-lab reproducibility may be interpreted similarly.

Sensitivity estimation by VAF ranges

Sensitivity is defined as the portion of known variants detected by each panel sequencing experiment among all the KPs targeted by the panel. As the list of KPs was not all inclusive, sensitivity was provided as an estimate. Fortunately, the number of known variants was usually high enough to provide an accurate estimate. For each sensitivity estimate, a 95% confidence interval was calculated by bootstrap resampling with the assumption that library replicates are not independent. Sensitivity estimates can be further stratified by the VAF of known variants. Four VAF ranges were adopted in this study, 1–2.5%, 2.5–5%, 5–10%, and 10–20%. For the spike-in sample, there was a single range around 5% as all AcroMetrix variants were spiked-in at a VAF of 5%.

FP rate estimation approaches

Three approaches were adopted to estimate false positive (FP) rate. In the CTR, a list of known negative (KN) positions was provided for sample A. This list covers over 50% of the CTR, enough for an accurate estimation of FP rate. By removing the positions overlapping with any known variants in sample B, a separate list of KN positions was generated for sample C. Any variant calls overlapping with the negative positions are flagged as FPs. FP rate is estimated as the ratio of FP calls out of every million KN positions. This is the first approach (coded as “viaKN”), applicable to sample A and sample C, but limited to the CTR.

By the second approach (coded as “B_low”), any calls in the normal sample B with VAF between 1 and 10% are marked as false positives (FPs). A pure diploid normal sample has variants at either 50% (heterozygous) or 100% (homozygous). Stochastic sampling and the influence of copy number variation lead to VAF deviation from those two values. As no extremely high copy number variation was observed in sample B, it is reasonable to assume that variant calls in sample B with VAF below 10% are FPs. Variant calls with VAF greater than 10% in any of other three sample B libraries at the same test sites will be excluded. The third approach (coded as “C_only”) leveraged sample mixture to differentiate FPs. Any variant in sample C must come from either sample A or sample B. If a variant is called in one sample C library, but not in any of four sample B libraries with VAF above 10% or any of four sample A libraries at the same test laboratory, it is an FP. This approach is more burdensome as all three samples and multiple libraries must be considered, but it does not rely on the assumption of sample B’s variant VAF range or the KNs in sample A. Furthermore, it cannot detect FPs due to any panel bias. Unlike “viaKN,” both “B_low” and “C_only” can be applied to estimate FP rate in the regions beyond the CTR.

Description of three sub-regions for each panel

The CTR was defined as the overlapping regions of four whole-exome sequencing panels that were used in the companion work²¹ to characterize known variants in sample A, NIST high-confidence regions, and exome coding regions. The CTR was then

refined by removing the low complexity regions. The CTR is completely within high-confidence exome coding regions (HC_CR) that is defined as the overlapping regions of NIST high-confidence regions and exome coding regions. By intersecting the CTR and HC_CR with the targeting region of each panel, three sub-regions for each panel can be defined for performance comparison: within the CTR, within the HC_CR but outside of the CTR, and the rest.

Evaluation of technical run-to-run variability in panel TMB estimates via simulation

The simulation aimed to reduce the variant density and reduce low VAF variants to better mimic a “real-world” clinical sample. While the high variant density and high percentage of variants with VAF below 5% in the test samples provided an advantage in estimating detection sensitivity with narrow confidence intervals, they presented obstacles here for the evaluation of TMB technical variability. The high variant density (equivalent to a TMB value about 1000) in the test samples will overshadow the effect of the FP calls and thus greatly underestimate TMB technical variability. The high percentage (over 75% in sample C) of variants with VAF below 5% may lead to an overestimate of TMB technical variability due to the variability in detecting these low VAF variants. The simulation was applied to the CTR. The resultant median VAF for low-frequency (<20%) KPs would be between 5 and 10%. Sample C was chosen for the simulation analysis as there are fewer uncharacterized true variants in sample C with VAF greater than 2.5% in comparison to sample A. All the known variants in sample B were excluded first from our KP list and from all the variant calling results (VCF files of library replicates) of each panel. The KPs with VAF < 5% or VAF > 40% in sample A were also removed. In theory, no KP with VAF < 2.5% or VAF > 20% in sample C was left. To mimic multiple samples with the constant TMB at a similar level with a clinical sample, we kept a fixed portion of KPs that resulted in the same number of KPs across the CTR for panel testing in the simulated samples. Specifically, based on their VAF in sample A, $k\%$ ($k = 0, 1, 2, 5, 10, 20$) of KPs were retained for VAF between 10 to 40% and $0.5 \times k\%$ of KPs were kept for VAF between 5 and 10%. Given each value, $(100 - k)\%$ of KPs were randomly excluded from CTR to generate a base-report-region for all the pan-cancer panels. Therefore, the variants not excluded from a VCF file were counted as the reported calls by the library of a pan-cancer panel in a simulation run. It is worth pointing out that the reported variants include FP calls and possibly uncharacterized true variants in addition to the retained KPs. Finally, 5000 rounds of simulation were performed for each based-report-region with a k value greater than 0.

For each one of 5000 base-report-region with a k value ($k > 0$), the reported variants were then counted for these above a chosen VAF threshold for each library replicate. The variance and mean across all technical replicates for each panel were calculated. Then, we computed the overall average of variance and mean over the 5000 runs of simulation. Finally, the CV for TMB were calculated as:

$$CV = \frac{\sqrt{\text{Overall Average of Variance}}}{\text{Overall Average of Mean Count}}$$

The mean TMB for each chosen k value was calculated as the overall average count of variants divided by the report region sizes of six panels. This CV is adopted to

measure the technical run-to-run variability and examine its dependence on TMB and VAF cutoff.

Model the deviation in TMB rate from WES by pan-cancer panels via TCGA dataset

Mutation Annotation Format (MAF) files of 10,295 cancer genomes across 33 cancer types were collected from The Cancer Genome Atlas (TCGA) project involving 2.9 million somatic mutations (MC3 Public MAF) for TMB measurement. The shared portion of TCGA coding sequence (CDS) region and the CTR with a size of 22,089,460 bases (55% of TCGA CDS) was used to calculate the TMB rate of each individual cancer sample. The TCGA-measured TMB rate was calculated as the number of mutations per million bases within the shared region and used as the baseline in this analysis. Then, given the TCGA somatic mutations as ground truth, the panel-measured TMB rate of each cancer sample was calculated by the number of mutations within the panel divided by the panel's region overlapping with the CTR and TCGA CDS in million base pairs. Two panels, ROC and TFS, were excluded from this analysis because of their small sizes.

For each pan-cancer panel, the mean squared deviation (MSD) was utilized to evaluate the difference between TCGA and panel TMB rates of the same group of cancer samples,

$$\text{MSD}(P_j) = \frac{\sum_{i=1}^n (\text{TCGA.TMB}_i - P_j(i, \text{TMB}_i, \text{Cov}_j))^2}{N}$$

where j represents each panel and i is an individual cancer sample. The function P_j calculates the TMB of sample i within the covered region (Cov_j) by panel j . N represents the number of samples in the group for the calculation of MSD. One outlier cancer genome TCGA-13-0889-01A-01W-0420-08 (with TCGA TMB rate of 33) was excluded from the calculation of MSD due to the abnormally huge difference between the TMB rates measured by TCGA WES and panels.

For a group of samples with TMB rates close to each other and thus close to their mean value, a modified MSD (MSD') is defined by substituting the TMB mean for the individual TCGA TMB in the equation above. Approximately, MSD equals to MSD' subtracting the variance of TCGA TMB in the sample group. Mathematically, MSD' equals to the sum of the variance of TMB by the panel and the squared difference in the mean TMB values between TCGA and the panel. Thus, MSD is broken down into two components of mean bias and variance (see Additional file 2: Supplementary Methods for details). The mean bias component is equal to the squared difference in the mean TMB values between TCGA and the panel. The variance component is equal to the variance of TMB by the panel reduced by the variance of TCGA TMB. This adjusted variance component is of particularly interest.

To better understand this adjusted panel TMB variance, we selected the TCGA samples (about 2400 samples) with TMB rates from 5 to 40 and sorted them by TMB. We then grouped each 100 samples and calculated the panel TMB mean value and the adjusted panel TMB variance. Multiplying it by the squared panel size in million bases, we converted the adjusted panel TMB variance to the mutation count level, which was then plotted against the product of the panel TMB mean and the panel size. A linear regression model was fitted over these two

variables from all the six panels that cover at least 250 kb in the CTR (Additional file 3: Fig. S5). A very high R^2 value (0.957) was observed along with a tight 95% confidence interval (1.28–1.34) for the slope of the regression line. This revealed a strong linear relationship between the adjust panel TMB variance and the panel TMB mean in mutation count. Finally, an intrinsic CV for each panel could be calculated as 1.15 divided by the square root of panel TMB mean in mutation count.

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s13059-021-02315-0>.

Additional file 1: Table S1. Detailed information for eight participating pan-cancer panels. **Table S2.** 95% confidence interval for reported sensitivity across VAF ranges for SNVs in the consensus targeted region (CTR). **Table S3.** 95% confidence interval for reported sensitivity in detecting known SNVs and other variants of expected VAF between 2.5% and 20%. **Table S4.** 95% confidence interval for reported sensitivity in detecting known SNVs within the CTR of expected VAF between 2.5% and 5% after applying the artificial VAF cutoff at 1.5%, 2.0% and 2.5%.

Table S5. Sensitivity across VAF ranges for all samples after applying the artificial VAF cutoff. **Table S6.** 95% confidence interval for reported sensitivity within the CTR or HC_CR (more specifically, in HC_CR beyond the CTR) in detecting known positives of expected VAF between 2.5% and 20%.

Additional file 2. Supplementary Methods.

Additional file 3: Fig. S1. Concordance of average cross-lab reproducibility and intra-lab reproducibility in Phred scale across VAF ranges for samples A and C. **Fig. S2.** Comparison of sensitivity of different variant types across VAF ranges. **Fig. S3.** Estimation of the false positive rate by three methods. **Fig. S4.** Low reproducibility of false positive calls across library replicates. **Fig. S5.** Scatter plot of panel TMB mean \times panel_size (x-axis) and adjusted panel TMB variance \times panel_size² (y-axis).

Additional file 4. Review history.

Acknowledgements

We thank our FDA colleagues Drs. Anand Pathak and Shashi Amur for providing useful feedback during study design and manuscript preparation.

Review history

The review history is available as Additional file 4.

Peer review information

Anahita Bishop was the primary editor of this article and managed its editorial process and peer review in collaboration with the rest of the editorial team.

Disclaimers

This research includes contributions from, and was reviewed by, the FDA and the NIH. This work has been approved for publication by these agencies, but does not necessarily reflect official agency policy. Certain commercial materials and equipment are identified in order to adequately specify experimental procedures. In no case does such identification imply recommendation or endorsement by the FDA or the NIH, nor does it imply that the items identified are necessarily the best available for purpose.

Authors' contributions

W.T., L.S., W.J., and J.X. conceived the project. B.G., DJJ Jr., J.C.W., W.T., L.S., W.J., and J.X. devised the experiments. N.N. and L. L. prepared the reference samples for cross-laboratory experiments. N.N., C. P-P., Z.Z., K.L., W.C., J.S.L., E.L., T.A.R., V.K.M., C.X., D.B., S.C., K.G., N.H., F.Q., K.W., C.A., S.H., M.J., W.Q., A.S., Y.W., G.Z., and S.Z. coordinated laboratory experiments and/or performed panel-specific variant calling analysis. G.C., H.A., A.B., J.B., B.B., D.J.B., A.C., C.R.C., D.C., J.C., J.C.C., D.J.C., E.C., A.d.P., D.D., A.K.E., J.F., A.G., S.G., C.H., X.H., B.H., J.H., K.H., J.H., J.K., B.K., P.L., Q.L., W.L., Y.L., S.L., C.M., N.M., R.M.-A., Q.M., P.A.M., D.M., C.P.P., A.R., P.R., R.R., B.S., E.S., R.S., R.S., J.C.S.-C., M.S., M.S.L., P.S., D.S., M.S., A.S., J.S., P.S., H.T., L.T., Y.T., V.J.T., D.T., E.V.G., S.V., K.W., V.W., J.Y., S.H.Y., M.Z., X.Z., and C.E.M. performed laboratory experiments. B.G., D.L., Y.Z., S.W., P.R.B., Y.Y., H.B., T.B., L.B., C.C., T.C., M.C., I.W.D., X.F., C.F., M.G., H.H., L.H., D.P.K., P.L., P.L., Z.L., Z.L., D.B.M., T.M., B.N., B.L.P., M.P., Q.S., T.S., S. T., D.T.-M., J.T.-M., B.T., N.T., C.W., J.W., Z.W., L.W., C.X., W.X., S.X., M.Y., Y.Z., and T.M. contributed to data analysis. B.G., D.L., Y.Z., and S.W. prepared the figures. R.K., B.G., D.L., W.J., L.S., and J.X. prepared the manuscript, with support from all co-authors. All authors read and approved the final manuscript.

Funding

All SEQC2 participants freely donated their time, reagents, and computing resources for the completion and analysis of this project. Part of this work was carried out with the support of the Intramural Research Program of the National Institutes of Health (to Mehdi Pirooznia), National Institute of Environmental Health Sciences (to Pierre Bushel), and National Library of Medicine (to Danielle Thierry-Mieg, Jean Thierry-Mieg, and Chunlin Xiao). Leming Shi and Yuanqing Zheng were supported by the National Key R&D Project of China (2018YFE0201600), the National Natural Science Foundation of China (31720103909), and Shanghai Municipal Science and Technology Major Project (2017SHZDX01).

Donald J. Johann, Jr. acknowledges the support by FDA BAA grant HHSF223201510172C. Timothy Mercer and Ira Deveson were supported by the National Health and Medical Research Council (NHMRC) of Australia grants APP1108254, APP1114016, and APP1173594 and Cancer Institute NSW Early Career Fellowship 2018/ECF013. This research has also been, in part, financially supported by the MEYS of the CR under the project CEITEC 2020 (LQ1601), by MH CR, grant No. (NV19-03-00091). Part of this work was carried out with the support of research infrastructure EATR IS-CZ, ID number LM2015064, funded by MEYS CR. Boris Tichy and Nikola Tom were supported by research infrastructure EATRIS-CZ, ID number LM2018133 funded by MEYS CR and MEYS CR project CEITEC 2020 (LQ1601). The contents of the published materials are solely the responsibility of the administering institution, a participating institution or individual authors, and they do not reflect the views of any funding body listed above.

Availability of data and materials

Important data items from the consortium oncopanel sequencing effort that we intend to disclose include the following: (1) NGS sequencing data [44] (in FASTQ or BAM format) from eight oncopanels for sample A, sample B, sample C, and the spike-in sample; (2) variant call results [45] (in VCF format) by the panel providers for these four samples; and (3) dependent files [45] (e.g., panel BED files, genome FASTA files) for each oncopanel.

Declarations

Ethics approval and consent to participate

Not Applicable.

Consent for publication

Not Applicable.

Competing interests

The authors declare the following potential competing financial interests:

Natalia Novorodovskaya, Katherine Wilkins, Anne Lucas, Scott Happe, and Carlos Pabon are all employees of Agilent Technologies, Inc. Agilent sample B DNA reference sample is a current product and sample A DNA and sample C DNA are potential products of Agilent Technologies, Inc.

Author details

¹Division of Bioinformatics and Biostatistics, National Center for Toxicological Research, US Food and Drug Administration, Jefferson, AR 72079, USA. ²Immuneering Corporation, One Broadway, 14th Floor, Cambridge, MA 02142, USA. ³Agilent Technologies, 11011 N Torrey Pines Rd, La Jolla, CA 92037, USA. ⁴Department of Information Science, University of Arkansas at Little Rock, 2801 S. Univ. Ave, Little Rock, AR 72204, USA. ⁵State Key Laboratory of Genetic Engineering, School of Life Sciences and Shanghai Cancer Hospital/Cancer Institute, Fudan University, Shanghai 200438, China. ⁶Agilent Technologies, 5301 Stevens Creek Blvd, Santa Clara, CA 95051, USA. ⁷Research and Development, Burning Rock Biotech, Shanghai 201114, China. ⁸Bioinformatics, Integrated DNA Technologies, Inc., 1710 Commercial Park, Coralville, IA 52241, USA. ⁹GeneTech, 8 Shengmingyuan Rd., Zhongguancun Life Science Park, Changping District, Beijing 100080, China. ¹⁰Illumina Inc., 5200 Illumina Way, San Diego, CA 92122, USA. ¹¹Research and Development, QIAGEN Sciences Inc., Frederick, MD 21703, USA. ¹²Market & Application Development Bioinformatics, Roche Sequencing Solutions Inc., 4300 Hacienda Dr, Pleasanton, CA 94588, USA. ¹³Thermo Fisher Scientific, 110 Miller Ave, Ann Arbor, MI 48104, USA. ¹⁴Clinical Diagnostic Division, Thermo Fisher Scientific, 46500 Kato Rd, Fremont, CA 94538, USA. ¹⁵Winthrop P Rockefeller Cancer Institute, University of Arkansas for Medical Sciences, 4301 W Markham St, Little Rock, AR 72205, USA. ¹⁶Departments of Medicine, Pathology, and Cancer Biology, College of Medicine and Life Sciences, University of Toledo Health Sciences Campus, 3000 Arlington Ave, Toledo, OH 43614, USA. ¹⁷National Institute of Environmental Health Sciences, Research Triangle Park, NC 27709, USA. ¹⁸Department of Immunology, Genomics and Microarray Core Facility, University of Texas Southwestern Medical Center, 5323 Harry Hine Blvd, Dallas, TX 75390, USA. ¹⁹Research and Development, Roche Sequencing Solutions Inc., 500 South Rosa Rd, Madison, WI 53719, USA. ²⁰Clinical Sequencing Division, Thermo Fisher Scientific, 180 Oyster Point Blvd, South San Francisco, CA 94080, USA. ²¹Marketing, Integrated DNA Technologies, Inc., 1710 Commercial Park, Coralville, IA 52241, USA. ²²Icahn Institute and Dept. of Genetics and Genomic Sciences Icahn School of Medicine at Mount Sinai, 1425 Madison Ave, New York, NY 10029, USA. ²³Greenwood Genetic Center, 106 Gregor Mendel Circle, Greenwood, SC 29646, USA. ²⁴Department of Pathology, Fudan University Shanghai Cancer Center, Fudan University, Shanghai 200032, China. ²⁵Department of Oncology, Shanghai Medical College, Fudan University, Shanghai 200032, China. ²⁶Institute of Pathology, Fudan University, Shanghai 200032, China. ²⁷JMP Life Sciences, SAS Institute Inc., Cary, NC 27519, USA. ²⁸Departments of Pathology and Pediatrics, University of Utah School of Medicine, Salt Lake City, UT 84108, USA. ²⁹R&D Genomics MPS, Institute for Clinical and Experimental Pathology ARUP Laboratories, 500 Chipeta Way, Salt Lake City, UT 84108, USA. ³⁰ResearchDx, Inc., 5 Mason, Irvine, CA 92618, USA. ³¹Department of Computer Science, Engineering and Physics, University of Michigan-Flint, Flint, MI 48502, USA. ³²Garvan Institute of Medical Research, Sydney, NSW, Australia. ³³St Vincent's Clinical School, University of New South Wales, Sydney, NSW 2010, Australia. ³⁴Department of Pathology, College of Medicine and Life Sciences, The University of Toledo, Toledo, OH 43614, USA. ³⁵Lucas County Coroner's Office, 2595 Arlington Ave., Toledo, OH 43614, USA. ³⁶Division of Gastroenterology and Hepatology, Mayo Clinic, Rochester, MN 55905, USA. ³⁷OmniSeq, Inc. 700 Ellicott St, Buffalo, NY 14203, USA. ³⁸Department of Physiology and Biophysics, Weill Cornell Medicine, Cornell University, New York, NY 10065, USA. ³⁹Stanford Genome Technology Center, Stanford University, Palo Alto, CA 94304, USA. ⁴⁰Division of Genetic and Molecular Toxicology, National Center for Toxicological Research, US Food and Drug Administration, Jefferson, AR 72079, USA. ⁴¹Fondazione Bruno Kessler, 38123 Trento, Italy. ⁴²Department of Medicine, College of Medicine and Life Sciences, The University of Toledo, Toledo, OH 43614, USA. ⁴³Institute of Medical and Molecular Genetics (INGEMM), Hospital Universitario La Paz, CIBERER Instituto de Salud Carlos III, 28046 Madrid, Spain. ⁴⁴EATRIS ERIC- European Infrastructure for Translational Medicine, De Boelelaan 1118, 1081 HZ Amsterdam, The Netherlands. ⁴⁵Kinghorn Centre for Clinical Genomics, Garvan Institute of

Medical Research, Sydney, NSW, Australia. ⁴⁶St Vincent's Clinical School, Faculty of Medicine, University of New South Wales, Sydney, NSW, Australia. ⁴⁷Cancer Genetics Inc, 201 Route 17 N, Meadows Office Building, Rutherford, NJ 07070, USA. ⁴⁸Institute for Personalized Cancer Therapy, MD Anderson Cancer Center, 6565 MD Anderson Blvd, Houston, TX 77030, USA. ⁴⁹Pharmaceutical Informatics Institute, College of Pharmaceutical Sciences, Zhejiang University, Hangzhou 310058, Zhejiang, China. ⁵⁰HK3 Lab, Milan, Italy. ⁵¹Molecular Laboratory, Prof. F. Raue, Im Weiher 12, Heidelberg, Germany. ⁵²Agilent Technologies, 1834 State Hwy 71 West, Cedar Creek, TX 78612, USA. ⁵³Department of Pathology, Strata Oncology, Inc., Ann Arbor, MI 48103, USA. ⁵⁴Human Genome Sequencing Center, Baylor College of Medicine, 1 Baylor Plaza, Houston, TX 77030, USA. ⁵⁵Massachusetts General Hospital, Harvard Medical School, Boston, MA 02114, USA. ⁵⁶NGS Products and Services, Integrated DNA Technologies, Inc., 1710 Commercial Park, Coralville, IA 52241, USA. ⁵⁷Molecular Genetics Laboratory, Molecular Diagnostics Division, London Health Sciences Centre, 800 Commissioners Rd E, London, Ontario N6A5W9, Canada. ⁵⁸Division of Anatomic Pathology, Mayo Clinic, 200 First Street SW, Rochester, MN 55905, USA. ⁵⁹Bioinformatics Research, Institute of Molecular Biotechnology, Boku University Vienna, Vienna, Austria. ⁶⁰Malopolska Centre of Biotechnology, Jagiellonian University, Krakow, Poland. ⁶¹Department of Biotechnology, Boku University, Vienna, Austria. ⁶²Institute of Medical and Molecular Genetics (INGEMM), Hospital Universitario La Paz, IdiPaz, CIBERER Instituto de Salud Carlos III, 28046 Madrid, Spain. ⁶³THACA, European Reference Network on Rare Congenital Malformations and Rare Intellectual Disability, European Commission, Lille, France. ⁶⁴Department of Pathology, National Cancer Center/National Clinical Research Center for Cancer/Cancer Hospital, Chinese Academy of Medical Sciences, No.17, Panjiayuan Nanli, Chaoyang District, Beijing 100021, China. ⁶⁵Center of Genome and Personalized Medicine, Institute of Cancer Stem Cell, Dalian Medical University, Dalian 116044, Liaoning, China. ⁶⁶Geneis, 5 Guangshun North St., Chaoyang District, Beijing 100102, China. ⁶⁷GeneSmile Ltd Co., Jiangsu Cancer Hospital, 42 Baiziting St., Xuanwu District, Nanjing 210009, Jiangsu, China. ⁶⁸Genycell Biotech España, Calle Garrido Atienza, 18320 Santa Fe, Granada, Spain. ⁶⁹CMINDS Research Center, Department of Electrical and Computer Engineering, College of Engineering, University of Massachusetts Lowell, Lowell, MA 01854, USA. ⁷⁰Department of Genetics, University of North Carolina, 250 Bell Tower Drive, Chapel Hill, NC 27599, USA. ⁷¹Accugenomics, Inc., 1410 Commonwealth Drive, Suite 105, Wilmington, NC 20403, USA. ⁷²Translational Research Laboratory, Belfer Center for Applied Cancer Science, Dana-Farber Cancer Institute, 360 Longwood Ave, Boston, MA 02215, USA. ⁷³Bioinformatics and Computational Biology Laboratory, National Heart Lung and Blood Institute, National Institutes of Health, Bethesda, MD 20892, USA. ⁷⁴Astrazeneca Pharmaceuticals, 35 Gatehouse Dr, Waltham, MA 02451, USA. ⁷⁵Department of Pathology and Laboratory Medicine, Western University, London, Ontario N6A3K7, Canada. ⁷⁶Institute for Molecular Medicine Finland (FIMM), Nordic EMBL Partnership for Molecular Medicine, HiLIFE Unit, Biomedicum Helsinki 2U (D302b), P.O. Box 20, (Tukholmankatu 8), FI-00014 University of Helsinki, Helsinki, Finland. ⁷⁷Laboratory for Molecular Genetics, Endocrine Practice, Im Weiher 12, 69121 Heidelberg, Germany. ⁷⁸Division of Systems Biology, National Center for Toxicological Research, US Food and Drug Administration, Jefferson, AR 72079, USA. ⁷⁹Center for Bioinformatics and Computational Biology, and the Institute of Biomedical Sciences, School of Life Sciences, East China Normal University, 500 Dongchuan Rd, Shanghai 200241, China. ⁸⁰National Centre for Cardiovascular Research (CNIC), Madrid, Spain. ⁸¹Primio Genes Biotechnology, Building C6-501, Biolake, No.666 Gaoxin Ave., East Lake High-tech Development Zone, Wuhan 430074, Hubei, China. ⁸²National Center for Biotechnology Information, National Library of Medicine, National Institutes of Health, 8600 Rockville Pike, Bethesda, MD 20894, USA. ⁸³Center of Molecular Medicine, Central European Institute of Technology, Masaryk University, Kamenice 5, 625 00 Brno, Czech Republic. ⁸⁴Center for Genomics, School of Medicine, Loma Linda University, Loma Linda, CA 92350, USA. ⁸⁵Division of Microbiology & Molecular Genetics, Department of Basic Sciences, School of Medicine, Loma Linda University, Loma Linda, CA 92350, USA. ⁸⁶Center for Individualized Medicine, Mayo Clinic, Scottsdale, AZ 85259, USA. ⁸⁷Department of Health Sciences, Mayo Clinic, Scottsdale, AZ 85259, USA. ⁸⁸Department of Molecular Pharmacology and Experimental Therapeutics, Mayo Clinic, Scottsdale, AZ 85259, USA. ⁸⁹College of Chemistry, Sichuan University, Chengdu 610064, Sichuan, China. ⁹⁰Science for Life Laboratory, Karolinska Institutet, Tomtebodavägen 23B, 171 65 Solna, Sweden. ⁹¹National Center for Biotechnology Information, National Library of Medicine, National Institutes of Health, 45 Center Drive, Bethesda, MD 20894, USA. ⁹²Department of Biostatistics, Columbia Mailman School of Public Health, 722 West 168th St., New York, NY 10032, USA. ⁹³Center for Genomic Sciences, LKS Faculty of Medicine, The University of Hong Kong, Hong Kong, SAR, China. ⁹⁴Clinical Laboratory, Burning Rock Biotech, Guangzhou 510300, Guangdong, China. ⁹⁵Geneplus, PKUCare Industrial Park, Changping District, Beijing 102206, China. ⁹⁶Australian Institute of Bioengineering and Nanotechnology, University of Queensland, Brisbane, QLD, Australia. ⁹⁷Genomics and Epigenetics Theme, Garvan Institute of Medical Research, Sydney, NSW, Australia. ⁹⁸Human Phenome Institute, Fudan University, Shanghai 201203, China. ⁹⁹Fudan-Gospel Joint Research Center for Precision Medicine, Fudan University, Shanghai 200438, China. ¹⁰⁰Q2 Solutions - EA Genomics, 5927 S Miami Blvd, Morrisville, NC 27560, USA.

Received: 28 July 2020 Accepted: 18 March 2021

Published online: 16 April 2021

References

- Lindeman NI, Cagle PT, Beasley MB, Chitale DA, Dacic S, Giaccone G, et al. Molecular testing guideline for selection of lung cancer patients for EGFR and ALK tyrosine kinase inhibitors: guideline from the College of American Pathologists, International Association for the Study of Lung Cancer, and Association for Molecular Pathology. *J Thorac Oncol*. 2013; 8(7):823–59. <https://doi.org/10.1097/JTO.0b013e318290868f>.
- Sosman JA, Kim KB, Schuchter L, Gonzalez R, Pavlick AC, Weber JS, McArthur GA, Hutson TE, Moschos SJ, Flaherty KT, Hersey P, Kefford R, Lawrence D, Puzanov I, Lewis KD, Amaravadi RK, Chmielowski B, Lawrence HJ, Shyr Y, Ye F, Li J, Nolop KB, Lee RJ, Joe AK, Ribas A. Survival in BRAF V600-mutant advanced melanoma treated with vemurafenib. *N Engl J Med*. 2012;366(8):707–14. <https://doi.org/10.1056/NEJMoa1112302>.
- Berger MF, Mardis ER. The emerging clinical relevance of genomics in cancer medicine. *Nat Rev Clin Oncol*. 2018;15(6):353–65. <https://doi.org/10.1038/s41571-018-0002-6>.
- Cordova-Delgado M, Pinto MP, Retamal IN, Muñoz-Medel M, Bravo ML, Fernández MF, Cisternas B, Mondaca S, Sanchez C, Galindo H, Nervi B, Ibáñez C, Acevedo F, Madrid J, Peña J, Koch E, Maturana MJ, Romero D, de la Jara N, Torres J,

- Espinoza M, Balmaceda C, Liao Y, Li Z, Freire M, Gárate-Calderón V, Cáceres J, Sepúlveda-Hermosilla G, Lizana R, Ramos L, Artigas R, Norero E, Crovari F, Armisen R, Corvalán AH, Owen GI, Garrido M. High proportion of potential candidates for immunotherapy in a Chilean cohort of gastric cancer patients: results of the FORCE1 study. *Cancers*. 2019;11(9):1275. <https://doi.org/10.3390/cancers11091275>.
5. Heeke AL, Pishvaian MJ, Lynce F, Xiu J, Brody JR, Chen W-J, et al. Prevalence of homologous recombination-related gene mutations across multiple cancer types. *JCO Precis Oncol*. 2018;2:PO.17.00286. <https://doi.org/10.1200/PO.17.00286>.
 6. Kacew AJ, Harris EJ, Lorch JH, Haddad RI, Chau NG, Rabinowits G, LeBoeuf NR, Schmults CD, Thakuria M, MacConaill LE, Hanna GJ. Chromosome 3q arm gain linked to immunotherapy response in advanced cutaneous squamous cell carcinoma. *Eur J Cancer*. 2019;113:1–9. <https://doi.org/10.1016/j.ejca.2019.03.004>.
 7. Hiley CT, Le Quesne J, Santis G, Sharpe R, de Castro DG, Middleton G, et al. Challenges in molecular testing in non-small-cell lung cancer patients with advanced disease. *Lancet*. 2016;388(10048):1002–11. [https://doi.org/10.1016/S0140-6736\(16\)31340-X](https://doi.org/10.1016/S0140-6736(16)31340-X).
 8. Martin P, Shiau CJ, Pasic M, Tsao M, Kamel-Reid S, Lin S, Tudor R, Cheng S, Higgins B, Burkes R, Ng M, Arif S, Ellis PM, Hubay S, Kuruville S, Laurie SA, Li J, Hwang D, Lau A, Shepherd FA, le LW, Leighl NB. Clinical impact of mutation fraction in epidermal growth factor receptor mutation positive NSCLC patients. *Br J Cancer*. 2016;114(6):616–22. <https://doi.org/10.1038/bjc.2016.22>.
 9. Groopman EE, Rasouly HM, Gharavi AG. Genomic medicine for kidney disease. *Nat Rev Nephrol*. 2018;14(2):83–104. <https://doi.org/10.1038/nrneph.2017.167>.
 10. Cheng DT, Mitchell TN, Zehir A, Shah RH, Benayed R, Syed A, Chandramohan R, Liu ZY, Won HH, Scott SN, Brannon AR, O'Reilly C, Sadowska J, Casanova J, Yannes A, Hechtman JF, Yao J, Song W, Ross DS, Oultache A, Dogan S, Borsu L, Hameed M, Nafa K, Arcila ME, Ladanyi M, Berger MF. Memorial Sloan Kettering-integrated mutation profiling of actionable cancer targets (MSK-IMPACT): a hybridization capture-based next-generation sequencing clinical assay for solid tumor molecular oncology. *J Mol Diagn*. 2015;17(3):251–64. <https://doi.org/10.1016/j.jmoldx.2014.12.006>.
 11. Gagan J, Van Allen EM. Next-generation sequencing to guide cancer therapy. *Genome Med*. 2015;7(1):80. <https://doi.org/10.1186/s13073-015-0203-x>.
 12. Samstein RM, Lee C-H, Shoushtari AN, Hellmann MD, Shen R, Janjigian YY, Barron DA, Zehir A, Jordan EJ, Omuro A, Kaley TJ, Kendall SM, Motzer RJ, Hakimi AA, Voss MH, Russo P, Rosenberg J, Iyer G, Bochner BH, Bajorin DF, al-Ahmadie HA, Chaff JE, Rudin CM, Riely GJ, Baxi S, Ho AL, Wong RJ, Pfister DG, Wolchok JD, Barker CA, Gutin PH, Brennan CW, Tabar V, Mellingerhoff IK, DeAngelis LM, Ariyan CE, Lee N, Tap WD, Gounder MM, D'Angelo SP, Saltz L, Stadler ZK, Scher HI, Baselga J, Razavi P, Klebanoff CA, Yaeger R, Segal NH, Ku GY, DeMatteo RP, Ladanyi M, Rizvi NA, Berger MF, Riaz N, Solit DB, Chan TA, Morris LGT. Tumor mutational load predicts survival after immunotherapy across multiple cancer types. *Nat Genet*. 2019;51(2):202–6. <https://doi.org/10.1038/s41588-018-0312-8>.
 13. Chalmers ZR, Connelly CF, Fabrizio D, Gay L, Ali SM, Ennis R, et al. Analysis of 100,000 human cancer genomes reveals the landscape of tumor mutational burden. *Genome Med*. 2017;9:34. <https://doi.org/10.1186/s13073-017-0424-2>.
 14. FoundationOne. FoundationOne CDx Technical Information. https://www.accessdata.fda.gov/cdrh_docs/pdf17/P170019C.pdf. Accessed 24 Feb. 2020.
 15. Omics Core by NantHealth, Inc. 510(k) Premarket Notification. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMN/pmn.cfm?ID=K190661>. Accessed 24 Mar 2020.
 16. PGDx elio tissue complete by Personal Genome Diagnostic, Inc. 510(k) Premarket Notification. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm?ID=K192063>. Accessed 13 Dec 2020.
 17. Lemery S, Keegan P, Pazdur R. First FDA approval agnostic of Cancer site — when a biomarker defines the indication. *N Engl J Med*. 2017;377(15):1409–12. <https://doi.org/10.1056/NEJMp1709968>.
 18. FDA approves third oncology drug that targets a key genetic driver of cancer, rather than a specific type of tumor. <https://www.fda.gov/news-events/press-announcements/fda-approves-third-oncology-drug-targets-key-genetic-driver-cancer-rather-specific-type-tumor>. Accessed 13 Mar 2020.
 19. Meric-Bernstam F, Brusco L, Shaw K, Horombe C, Kopetz S, Davies MA, Routbort M, Piha-Paul SA, Janku F, Ueno N, Hong D, de Groot J, Ravi V, Li Y, Luthra R, Patel K, Broadus R, Mendelsohn J, Mills GB. Feasibility of large-scale genomic testing to facilitate enrollment onto Genomically matched clinical trials. *J Clin Oncol*. 2015;33(25):2753–62. <https://doi.org/10.1200/JCO.2014.60.4165>.
 20. Gray SW, Hicks-Courant K, Cronin A, Rollins BJ, Weeks JC. Physicians' attitudes about multiplex tumor genomic testing. *J Clin Oncol*. 2014;32(13):1317–23. <https://doi.org/10.1200/JCO.2013.52.4298>.
 21. Jameson GS, Petricoin EF, Sachdev J, Liotta LA, Loesch DM, Anthony SP, Chadha MK, Wulfschlegel JD, Gallagher RI, Reeder KA, Pierobon M, Fulk MR, Cantafio NA, Dunetz B, Mikrut WD, von Hoff DD, Robert NJ. A pilot study utilizing multi-omic molecular profiling to find potential targets and select individualized treatments for patients with previously treated metastatic breast cancer. *Breast Cancer Res Treat*. 2014;147(3):579–88. <https://doi.org/10.1007/s10549-014-3117-1>.
 22. Tsongalis GJ, Peterson JD, de AFB, Tunkey CD, Gallagher TL, Strausbaugh LD, et al. Routine use of the ion torrent AmpliSeq™ Cancer hotspot panel for identification of clinically actionable somatic mutations. *Clin Chem Lab Med*. 2013;52(5):707–14. <https://doi.org/10.1515/cclm-2013-0883>.
 23. Garofalo A, Sholl L, Reardon B, Taylor-Weiner A, Amin-Mansour A, Miao D, Liu D, Oliver N, MacConaill L, Ducar M, Rojas-Rudilla V, Giannakis M, Ghazani A, Gray S, Janne P, Garber J, Joffe S, Lindeman N, Wagle N, Garraway LA, van Allen EM. The impact of tumor profiling approaches and genomic data strategies for cancer precision medicine. *Genome Med*. 2016;8(1):79. <https://doi.org/10.1186/s13073-016-0333-9>.
 24. Kuderer NM, Burton KA, Blau S, Rose AL, Parker S, Lyman GH, Blau CA. Comparison of 2 commercially available next-generation sequencing platforms in oncology. *JAMA Oncol*. 2017;3(7):996–8. <https://doi.org/10.1001/jamaoncol.2016.4983>.
 25. Garcia EP, Minkovskiy A, Jia Y, Ducar MD, Shivdasani P, Gong X, Ligon AH, Sholl LM, Kuo FC, MacConaill LE, Lindeman NI, Dong F. Validation of OncoPanel: a targeted next-generation sequencing assay for the detection of somatic variants in cancer. *Arch Pathol Lab Med*. 2017;141(6):751–8. <https://doi.org/10.5858/arpa.2016-0527-OA>.
 26. Jones WD. SEQC2 Oncopanel sequencing working group. A verified genomic reference sample for assessing performance of cancer panels detecting small variants of low allele frequency. *Genome Biol*. <https://doi.org/10.1186/s13059-021-02316-z>.

27. MAQC Consortium. The MicroArray quality control (MAQC) project shows inter- and intraplatform reproducibility of gene expression measurements. *Nat Biotechnol.* 2006;24(9):1151–61. <https://doi.org/10.1038/nbt1239>.
28. SEQC/MAQC-III Consortium. A comprehensive assessment of RNA-seq accuracy, reproducibility and information content by the sequencing quality control consortium. *Nat Biotechnol.* 2014;32(9):903–14. <https://doi.org/10.1038/nbt.2957>.
29. MAQC Consortium, Shi L, Campbell G, Jones WD, Campagne F, Wen Z, et al. The MicroArray Quality Control (MAQC)-II study of common practices for the development and validation of microarray-based predictive models. *Nat Biotechnol.* 2010;28(8):827–38. <https://doi.org/10.1038/nbt.1665>.
30. Shi L, Perkins RG, Fang H, Tong W. Reproducible and reliable microarray results through quality control: good laboratory proficiency and appropriate data analysis practices are essential. *Curr Opin Biotechnol.* 2008;19(1):10–8. <https://doi.org/10.1016/j.copbio.2007.11.003>.
31. Xiao W, SEQC2 Somatic Mutation Working Group. Achieving reproducibility and accuracy in cancer mutation detection with whole-genome and whole-exome sequencing. *bioRxiv.* 2019. <https://doi.org/10.1101/626440>. Accessed 24 Feb 2020.
32. Fang LT, SEQC2 Somatic Mutation Working Group. Establishing reference samples for detection of somatic mutations and germline variants with NGS technologies. *bioRxiv.* 2019. <https://doi.org/10.1101/625624>. Accessed 24 Feb 2020.
33. Sha D, Jin Z, Budczies J, Kluck K, Stenzinger A, Sinicrope FA. Tumor mutational burden as a predictive biomarker in solid tumors. *Cancer Discov.* 2020;10(12):1808–25. <https://doi.org/10.1158/2159-8290.CD-20-0522>.
34. Novoradovskaya N, Whitfield ML, Basehore LS, Novoradovsky A, Pesich R, Usary J, Karaca M, Wong WK, Aprelikova O, Fero M, Perou CM, Botstein D, Braman J. Universal reference RNA as a standard for microarray experiments. *BMC Genomics.* 2004;5(1):20. <https://doi.org/10.1186/1471-2164-5-20>.
35. Thermo Scientific. AcroMetrix Oncology Hotspot Control Package Insert. <https://www.thermofisher.com/document-connect/document-connect.html?url=https%3A%2F%2Fassets.thermofisher.com%2F%2FAssets%2FCDD%2Fmanuals%2FFMAN0010820-AMX-Oncology-Hotspot-Ctrl-EN.pdf&title=QWNYb01ldHJpeCBPbmNvbG9neSBib3RzcG90ENVbnRyb2wgUGFja2FnZSBjbNlcnQgW0VOXQ==>. Accessed 24 Apr 2019.
36. Ellrott K, Bailey MH, Saksena G, Covington KR, Kandath C, Stewart C, et al. Scalable open science approach for mutation calling of tumor exomes using multiple genomic pipelines. *Cell Syst.* 2018;6(3):271–81.e7. <https://doi.org/10.1016/j.cels.2018.03.002>.
37. Campbell PJ, Getz G, Korbel JO, Stuart JM, Jennings JL, Stein LD, et al. Pan-cancer analysis of whole genomes. *Nature.* 2020;578(7793):82–93. <https://doi.org/10.1038/s41586-020-1969-6>.
38. Sims D, Sudbery I, Illott NE, Heger A, Ponting CP. Sequencing depth and coverage: key considerations in genomic analyses. *Nat Rev Genet.* 2014;15(2):121–32. <https://doi.org/10.1038/nrg3642>.
39. Zhang Y, SEQC2 OncoPanel Sequencing Working Group. Deep oncopanel sequencing reveals fixation time- and within block position-dependent quality degradation in FFPE processed samples. *bioRxiv.* 2021. <https://doi.org/10.1101/2021.04.06.438687>. Accessed 8 Apr 2021.
40. Willey JC, Morrison T, Austerhammer B, Crawford EL, Craig DJ, Blomquist TM, et al. Advancing quality-control for NGS measurement of actionable mutations in circulating tumor DNA. *bioRxiv.* 2021. <https://doi.org/10.1101/2021.04.06.438497>. Accessed 8 Apr 2021.
41. Allgäuer M, Budczies J, Christopoulos P, Endris V, Lier A, Rempel E, Volckmar AL, Kirchner M, von Winterfeld M, Leichsenring J, Neumann O, Fröhling S, Penzel R, Thomas M, Schirmacher P, Stenzinger A. Implementing tumor mutational burden (TMB) analysis in routine diagnostics—a primer for molecular pathologists and clinicians. *Transl Lung Cancer Res.* 2018;7(6):703–715–15. <https://doi.org/10.21037/tlcr.2018.08.14>.
42. Buchhalter I, Rempel E, Endris V, Allgäuer M, Neumann O, Volckmar A-L, Kirchner M, Leichsenring J, Lier A, von Winterfeld M, Penzel R, Christopoulos P, Thomas M, Fröhling S, Schirmacher P, Budczies J, Stenzinger A. Size matters: dissecting key parameters for panel-based tumor mutational burden analysis. *Int J Cancer.* 2019;144(4):848–58. <https://doi.org/10.1002/ijc.31878>.
43. Budczies J, Allgäuer M, Litchfield K, Rempel E, Christopoulos P, Kazdal D, Endris V, Thomas M, Fröhling S, Peters S, Swanton C, Schirmacher P, Stenzinger A. Optimizing panel-based tumor mutational burden (TMB) measurement. *Ann Oncol.* 2019;30(9):1496–506. <https://doi.org/10.1093/annonc/mdz205>.
44. Gong B, Xu J. SEQC2 Onco-panel Sequencing Working Group - PanCancer panel Study. BioProject PRJNA677997. NCBI. 2021. <https://www.ncbi.nlm.nih.gov/bioproject/PRJNA677997>. Accessed 26 Feb 2021.
45. Gong B, Xu J. SEQC2 Onco-panel Sequencing Working Group - PanCancer panel Study. figshare. 2021. https://figshare.com/projects/SEQC2_Onco-panel_Sequencing_Working_Group_-_PanCancer_panel_Study/94520. Accessed 25 Feb 2021.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

